

LARGE REFINERY SAVES \$49,200 IN REGULATORY COMPLIANCE COSTS WITH 8714 METER VERIFICATION DIAGNOSTIC

Application

Well water flows

Customer

A large refinery located in the United States

Challenge

As part of the state Environmental Protection Agency (EPA) requirements, this customer is required to accurately account for the amount of well water they extract from the ground. This is accomplished by using magnetic flow meter technology to measure and totalize the amount of water consumption from each of their seven water wells. In order to confirm that the meters are measuring accurately, the state EPA requires that the meter calibration be verified on an annual basis. In order to satisfy the EPA requirements, this customer would have to send their meters to a flow lab each year to ensure that the calibration had not changed, thus verifying the usage being reported to the state was accurate. This required two Instrument Technicians to remove each meter and ship it to a flow lab for calibration, which increased maintenance costs. A spare meter was required to be installed while the flow meter was being calibrated in the flow lab. Since some of the wells were the same size, a total of four spare meters were used on a rotating basis, which increased inventory costs.

Results

- Reduced maintenance costs
- Improved ease of regulatory compliance
- Reduced spare inventory requirements

By using 8714 meter verification diagnostics, this customer satisfied their state EPA meter verification requirements without having to remove the flow meter from the line or maintain any spare meters.



The Rosemount[™] 8700 Flow meter



LARGE REFINERY SAVES \$49,200 IN REGULATORY COMPLIANCE COSTS WITH 8714 METER VERIFICATION DIAGNOSTIC


Solution

In order to reduce the costs associated with meter verification, it was recommended that this customer use Rosemount 8700 flow meters with in-situ meter verification diagnostic capabilities. The 8714 meter verification diagnostic performed a full meter verification of the sensor by taking a baseline signature of the sensor when it was first installed.

Running the 8714 meter verification compared the original signature to the current measurements. Since the values were within the user set parameters, the flow meter was within specification. The 8714 meter verification diagnostic ran in minutes, without the need to remove the flow meter from the line.

By using 8714 meter verification diagnostics, this customer satisfied their state EPA meter verification requirements without having to remove the flow meter from the line or maintain any spare meters. By not having to send the meters out for calibration, the customer eliminated flow lab costs, maintenance costs, and shipping costs, resulting in a \$49,200 cost savings.

A meter verification report can be printed from AMS[®] Suite: Intelligent Device Manager. Alternatively, the calibration values can be taken from the local operator interface or field communicator and written into the report by hand. The report can then be submitted to any regulatory agency to certify that the meter calibration has not changed, satisfying verification requirements.



The Emerson logo is a trademark and service mark of Emerson Electric Co. Brand logotype are registered trademarks of one of the Emerson family of companies. All other marks are the property of their respective owners. © 2024 Emerson Electric Co. All rights reserved.

For more information, visit Oil & Gas

Emerson.com/OilandGas Magmeter Learn About Emerson.com/RosemountMagneticFlow

00830-0700-4662, Rev AC

a lavalar sales	the second se
Overview Critical Informational Diagnostics 8714i Report 8714i Report 8714i Calibration Verification Report	
	Calibration Conditions: 🖓 Internal 🔽 External
Customer:	
Tag	Test Conditions: No Flow, Full Pipe
Flowmeter Information and Configuration	
Tag	PV URV 75.00 gal/min
Calibration Number 1035006610326005	PV LRV 0.00 gal/min
ine Size 1.50 in	PV Damping 2.00 s
Transmitter Calibration Verification Results	Flowtube Sensor Calibration Verification Results
	Flowtube Deviation % : 0.490468
Simulated Velocity Actual Velocity Dev	Tube Calibration Test: Pass -
30.000000 30.016661 0	106 Pass _ Col Circuit Test. Pass _
	Electrode Circuit Test (if applicable) Pass
Summary of Calibration Venification Results	
Verification Results: The result of the flowmet	er venification test is: Pass 👱
Verification Criteria: This meter was verified to	be functioning within % of deviation from the original test paramet
Signed	Date:

