

Subject: Micro Motion CDM Software Release 2.20

Overview

This bulletin documents the release of new software for the CDM. The version of this software is 2.20. Weights and Measures version 1.1.

Products Affected

CDM Analog (Transmitter Output Options C and D) and TPS (Transmitter Output Option B) versions

NAMUR NE 53 Rating

This release is rated at NAMUR NE 53 Level 2 with a NAMUR NE 53 version of 01.02.02.

Symptoms

The following symptoms may appear with CDM:

- The meter would not start up when power cycled at or above 60°C.
- The API reference temperature was limited to lower values than specified by the 2004 standard.
- When using API table E, if both the density and temperature were out of limits, only the temperature out of limits alert was displayed.

Resolution

- All of the above issues have been corrected with the release of version 2.20.
- No known safety issues remain open.
- For firmware updates to a unit in the field that is experiencing any of these issues, please contact Micro Motion service in your area.

Additional Features

No new features were added for this release.

Firmware Checksum: 0x4D9E3DC

Contact Information

Product information is available on the internet at: www.emerson.com.

Customer Service Phone Numbers:

Micro Motion USA	1-800-522-6277
Micro Motion Europe	31 (0) 318 495 555
Micro Motion Asia	65 6777-8211
Micro Motion UK	44 0870 240 1978
Micro Motion Japan	81 3 5769-6803

At Micro Motion we strive to stay on the leading edge of Coriolis technology. Upgrading to our latest MVD technology can pay for itself in a very short time period by improving your process measurement, improving diagnostic coverage, and improving the bottom line results of your process.

Thank you for your continued support.

© 2020 Micro Motion, Inc. All rights reserved.

ELITE, ProLink, and the Micro Motion logo are registered trademarks, and MVD and MVD Direct Connect are trademarks of Micro Motion, Inc., Boulder, Colorado. The Emerson logo is a trademark of Emerson Electric Co. All other trademarks are property of their respective owners.

Micro Motion supplies this publication for informational purposes only. While every effort has been made to ensure accuracy, this publication is not intended to make performance claims or process recommendations. Micro Motion does not warrant, guarantee, or assume any legal liability for the accuracy, completeness, timeliness, reliability, or usefulness of any information, product, or process described herein. For actual product information and recommendations, please contact your local Micro Motion representative.